










# SBio HBsAg Test **DIPSTICK**

Rapid one step test for the detection of HBsAg in serum, plasma

REF	90202025
▽	25 T

 Store at 4 to 30 °C	 Manufacturer	<b>LOT</b> Batch Number	<b>DIPSTICK</b> Dipstick	<b>HBsAg</b> Rapid one step test for detection of HBsAg in serum, plasma	 Xn H302, H332 R22 S23-S24/25 S63 Harmful if swallowed. Do not breathe vapour. If swallowed, seek medical advice immediately and show this container or label. Avoid release to the environment. Refer to special instructions.
 Use by (Last day of stated month)	 Consult Instructions for use	<b>IVD</b> In vitro Diagnostic Medical Device	 This side up		
 Date of Manufacture	<b>REF</b> Catalogue Number	 Contains sufficient for <n> tests	 Do not reuse		

## INTRODUCTION

SBio HBsAg Test is a rapid, qualitative, two site sandwich immunoassay for the detection of Hepatitis B surface antigen, a marker for Hepatitis B infections, in serum/plasma specimen.

## SUMMARY

Blood containing the Hepatitis B Virus (HBV) is potentially infectious. Hepatitis B Surface Antigen (HBsAg), earlier known as Australia Antigen, is among the first serological markers that circulate in the blood of infected persons even two to three weeks prior to the appearance of clinical symptoms. The levels of HBsAg are especially elevated during the symptomatic phase and decline thereafter. Detection of HBV using HBsAg as the marker to screen blood donors is essential to reduce the risk of transmission of Hepatitis B by blood transfusion. HBsAg detection is also useful for screening high risk groups for HBV and for differential diagnosis of Hepatitis infection. SBio HBsAg Test detects the presence of HBsAg in serum/plasma specimens, qualitatively, at concentrations as low as 0.5 ng/ml.

## PRINCIPLE

SBio HBsAg Test utilizes the principle of immunochromatography, a unique two site immunoassay on a membrane. As the test sample flows through the membrane assembly of the test dipstick, the colored monoclonal anti-HBsAg-colloidal gold conjugate complexes with the HBsAg in the sample. This complex moves further on the membrane to the test region where it is immobilized by another monoclonal anti-HBsAg antiserum coated on the membrane leading to formation of a pink-purple colored band which confirms a positive test result. Absence of this colored band in the test region indicates a negative test result. The unreacted conjugate and unbound complex if any move further on the membrane and are subsequently immobilized by the anti-rabbit antiserum coated on the membrane at the control region, forming a pink-purple band. This control band serves to validate the test results.

## REAGENTS AND MATERIALS SUPPLIED

Each kit contains:

A. Individual pouches each containing a:

- DIPSTICK**: Membrane test assembly pre-dipsensed with monoclonal anti-HBsAg antiserum-colloidal gold conjugate, rabbit IgG colloidal gold conjugate and monoclonal anti-HBsAg antiserum and anti-rabbit antiserum coated at the respective regions.
- Desiccant pouch.

B. Package insert.

## STORAGE AND STABILITY

The sealed pouches in the test kit may be stored between 4 to 30°C till the duration of the shelf life as indicated on the pouch. DO NOT FREEZE.

## NOTE

- For in vitro diagnostic use only. NOT FOR MEDICINAL USE. For professional use.
- Do not use beyond expiry date.
- Do not reuse the test dipstick.
- Read the instructions carefully before performing the test.
- Handle all specimens as potentially infectious.
- Follow standard biosafety guidelines for handling and disposal of potentially infective material.
- If desiccant colour at the point of opening the pouch has turned from blue to pink, another test dipstick must be run.

## SPECIMEN COLLECTION AND PREPARATION

No special preparation of the patient is necessary prior to specimen collection by approved techniques. Though fresh serum/plasma is preferable, serum/plasma specimens may be stored 2 to 8°C for upto 24 hours, in case of delay in testing. Do not use haemolysed, turbid or contaminated samples. Turbid samples should be centrifuged and clear supernatant must be used for testing.

## TESTING PROCEDURE AND INTERPRETATION OF RESULTS

- Collect serum/plasma specimen in a clean test tube. Ensure that only sufficient quantity of the specimen is collected to allow submerging the red area of the dipstick (about 1 cm high).
- Bring the sealed pouch to room temperature, open the pouch and remove the dipstick (taking care not to touch the membrane area), and desiccant pouch. Check the color of the desiccant. It should be blue, if it has turned colorless or pink, discard the dipstick and use another dipstick. **Once opened, the dipstick must be used immediately.**
- Dip the dipstick in the serum/plasma specimen **submerging only the red area.**
- At the end of fifteen minutes read the results as follows:

**NEGATIVE:** A colored band appears on the dipstick.

**POSITIVE:** Two distinct colored bands appear on the dipstick.

**INVALID:** The test should be considered invalid if no colored bands appears on the dipstick. The test should also be considered invalid if a colored band appears only at the test region and not at the control region. And in such cases repeat the test with a new dipstick ensuring that the test procedure has been followed accurately.



## PERFORMANCE CHARACTERISTICS

### Internal Evaluation-I

In an in-house study, the performance of SBio HBsAg Test was evaluated using a panel of fifty known positives (of varying reactivity) and two hundred known negative specimens in comparison to two licensed ELISA kits ELISA-I & ELISA-II.

The results of the evaluation are as follows:

	TOTAL	SBio HBsAg Test	ELISA I	ELISA II
Number of specimens tested	250	250	250	250
Number of positives	50	50	50	50
Number of negatives	200	200	200	200

Based on this evaluation:

Sensitivity of SBio HBsAg Test: 100%. Specificity of SBio HBsAg Test: 100%.

### Internal Evaluation-II

SBio HBsAg Test was evaluated with a serial dilution of known concentration of HBsAg positive sample. It was observed that SBio HBsAg Test was able to detect all the dilutions with HBsAg concentration of  $\geq 0.5$  ng/ml.

Therefore the detection limit of SBio HBsAg Test is  $\geq 0.5$  ng/ml.

With a low titre performance panel (PHA 104) from Boston Biomedica Inc., USA, SBio HBsAg Test showed ( $\pm$ ) reactivity with a sample that contained as low as 0.3 ng/ml of HBsAg. In the same panel, with another sample of 0.6 ng/ml, SBio HBsAg Test showed (+) reactivity.

### Independent External Evaluation

In another independent study, the performance of SBio HBsAg Test was evaluated using a panel of 50 samples; 20 positives & 30 negatives, in comparison with commercially available Immunochromatographic Test (ICT), Enzyme Immunoassay (EIA) and Microparticle Enzyme Immunoassay (MEIA). The results of the evaluation are as follows:

	Total	SBio HBsAg Test	ICT	EIA	MEIA
Number of specimen tested	50	50	50	50	50
No. of Positives	20	19	18	20	20
No. of Negatives	30	31	32	30	30

The above study indicates good correlation of the results of SBio HBsAg Test with that of EIA & MEIA.

### LIMITATIONS OF THE TEST

- Though SBio HBsAg Test is a reliable screening assay, it should not be used as a sole criterion for diagnosis of HBV infection.
- Interference due to heterophile antibodies, rheumatoid factors and other nonanalyte substances in patients serum, capable of binding antibodies multivalently and providing erroneous analyte detection in immunoassays, has been reported in various studies.

Though SBio HBsAg Test uses sufficient amounts of HETEROPHILIC BLOCKING REAGENT (HBR) to inhibit the majority of this interference; nevertheless, some samples with high titres may still express clinically important assay interference. Both laboratory professionals and clinicians must be vigilant to this possibility of antibody interference. Results that appear to be internally inconsistent or incompatible with the clinical presentation should invoke suspicion of the presence of endogenous artifacts and lead to appropriate in vitro investigative action.

- Do not compare the intensity of test lines and the control lines to judge the concentration of HBsAg in the test specimen.
- Since various tests of HBsAg differ in their performance characteristics and antibody composition, their reactivity patterns may differ.
- Testing of pooled samples is not recommended.
- The membrane is laminated with an adhesive tape to prevent surface evaporation. Air pockets or patches may appear, which do not interfere with the test results. Presence of a band at the test region, even if low in intensity or formation, is a positive result.
- Most positive results develop within 15 minutes. However, certain sera sample may take a longer time to flow. Therefore, negatives should be confirmed only at 30 minutes. Do not read results after 30 minutes.
- HBsAg is coded for by the S gene, and the common antigenic epitopes of all subtypes of HBsAg are found in the same 'a' determinant. The monoclonal antibodies used in SBio HBsAg Test are directed against this 'a' determinant so that all subtypes of HBsAg can be detected. However, a few patients infected with HBV may show negative for HBsAg inspite of a positive test for HBV-DNA or HBV polymerase chain reaction. These rare cases are due to antigenically divergent variants. Therefore, the existence of such variants should be considered before taking clinical decisions.
- As with all diagnostic tests a definitive clinical diagnosis should not be based on the result of a single test, but should only be made by the physician after all clinical and laboratory findings have been evaluated.

### REFERENCES:

- (1) Kim, C.Y., Tillis, J.G.. 1973, J. Clin. Invest. 52, 1176-1186. (2) Magnus, L.O. et al. 1975, J. Am. Med. Assoc. 231, 356-359. (3) WHO, 1977. Technical Report Series, 602. (4) Caldwell, C.W., Barret, J.T. 1977. Clin. Chim. Acta. 81, 305. (5) Wolters, G. et al. 1977. J. Infect. Dis., 136, 311. (6) Rubin, E. 1979. Federation Proceedings. 38, 2665-2673. (7) Darrel L. et al. 1982. J. Biol. Chem., 257, 10414-10420. (8) Kuhn, P. et al. 1985 Lancet, May 25, 1222-1223. (9) Carman WF et al. Vaccine-induced escape mutant of hepatitis B virus. Lancet, 1990; 336:325029. (10) Koyanagi T et al. Analysis of HBs antigen negative variant of Hepatitis B virus: Unique substitutions, Glu 129 to Asp and Gly 145 to Ala in the surface antigen gene. MedSci Monit 2000;6(6):1165-1169.

Manufactured by:

**Zephyr Biomedicals**

A Division of Tulip Diagnostics (P) Ltd.

M 46-47, Phase III B, Verna Industrial Estate,  
Verna, Goa - 403 722, INDIA.

Regd. Office: Gitanjali, Tulip Block, Dr. Antonio Do Rego Bagh,  
Alto Santacruz, Bambolim Complex P.O., Goa - 403 202, INDIA.

Manufactured for:

**Singapore Biosciences PTE Ltd.**

11 Yishun Street 51, #04-23, The Criterion,  
Singapore 767971

EC REP

CMC Medical Devices & Drugs S.L., C/ Horacio Lengo No. 18, CP 29006, Malaga, Spain